

INSTAND e.V.

Gesellschaft zur Förderung der Qualitätssicherung in medizinischen Laboratorien e. V.
40223 Düsseldorf Ubier - Str. 20 , 40093 Postfach 250211, Germany

Certificate

Virus genome detection June 2018

Deadline of EQAS : 29.6.2018

Mgr. Petra Svastova (participant 29506 E)
GeneProof a.s.
Videnska 119
61900 Brno
Czech Republic


This is to certify that the above mentioned participant has s u c c e s s f u l l y taken part in EQAS June 2018 with the following analyses:


PCR-/ NAT - HSV 1 / HSV 2 (363) - Quantitative genome detection of HSV 1 (5) R-B3 **
PCR-/ NAT - HSV 1 / HSV 2 (363) - Qualitative genome detection of HSV 1 (10) R-B3 **
PCR-/ NAT - HSV 1 / HSV 2 (363) - Quantitative genome detection of HSV 2 (15) R-B3 **
PCR-/ NAT - HSV 1 / HSV 2 (363) - Qualitative genome detection of HSV 2 (20) R-B3 **
PCR-/ NAT - VZV (366) - Quantitative detection of VZV (10) R-B3 **
PCR-/ NAT - VZV (366) - Qualitative detection of VZV (20) R-B3 **
PCR-/ NAT - Enteroviruses (372) - Qualitative detection of Enteroviruses (20) R-B3 **
PCR-/ NAT - HPV (373) - Typing of "High Risk" - HPV (without detection of "Low Risk" - HPV) (35) R-B3 **

* valid for 6 months ** valid for 12 months *** valid for 2 years

R-B2: RilibÄK-parameter subject to B2 R-B3: RilibÄK-parameter subject to B3




Prof. Dr. M. Spannagel
(Chairman of the reference institution)


Prof. Dr. H. Zeichhardt
(Expert)

Düsseldorf, 8.8.2018

June 2018

Mgr. Petra Svastova (participant 29506 E)
GeneProof a.s.
Videnska 119
61900 Brno
Czech Republic

Valid results:
PCR-/ NAT - Enteroviruses (372)

sample test	qual. result
372066 Quantitative detection of Enteroviruses (10)	no evaluation
372067 Quantitative detection of Enteroviruses (10)	no evaluation
372068 Quantitative detection of Enteroviruses (10)	no evaluation
372069 Quantitative detection of Enteroviruses (10)	no evaluation
372066 Qualitative detection of Enteroviruses (20)	positive
372067 Qualitative detection of Enteroviruses (20)	positive
372068 Qualitative detection of Enteroviruses (20)	negative
372069 Qualitative detection of Enteroviruses (20)	positive
372066 Type identification of Enteroviruses (40)	type=ENT68
372067 Type identification of Enteroviruses (40)	type=EV7
372068 Type identification of Enteroviruses (40)	negative
372069 Type identification of Enteroviruses (40)	type=EV7

Your results:

sample test - method / reagent	evaluation of your results	your qual. result
Qualitative detection of Enteroviruses (20)		
372066 GeneProof Enterovirus PCR Kit	correct	positive
372067 GeneProof Enterovirus PCR Kit	correct	positive
372068 GeneProof Enterovirus PCR Kit	correct	negative
372069 GeneProof Enterovirus PCR Kit	+ correct	positive