


<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 <small>Quality Control for Molecular Diagnostics</small>	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### Intended Results / Panel Composition

Sample Code	Sample Content	Matrix	Sample Relationships <sup>[1]</sup>	Detection Frequency <sup>[2]</sup>	Sample Status <sup>[3]</sup>	Consensus (IU/ml) <sup>[4]</sup>		Range <sup>[5]</sup>
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-01	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	4.388	18	3.753 - 5.118
HHV6DNA19C1-02	HHV6 Negative	Plasma		Negative	CORE	N/A	N/A	N/A
HHV6DNA19C1-03	HHV6 Type A	Plasma		Frequently Detected	CORE	4.775	18	3.773 - 5.592
HHV6DNA19C1-04	HHV6 Type B	Plasma		Frequently Detected	EDUCATIONAL	2.901	17	2.107 - 3.518
HHV6DNA19C1-05	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	4.151	18	3.138 - 4.526

[1] **Sample Relationships:** Indicates the relationships of the samples within this challenge. The highest titre member of dilution series DS1 is indicated by DS1\_1 and further members of the series as DS1\_2, DS1\_3 etc. in order of reducing titre. Additional dilution series are indicated by DS2 (e.g. DS2\_1, DS2\_2 etc.), DS3 (e.g. DS3\_1, DS3\_2 etc.). If one duplicate pair is present this is indicated by 'D1'. Further duplicate pairs are indicated by 'D2', 'D3' etc.

[2] **Detection Frequency:** To aid qualitative analysis each panel member is assigned a frequency of detection. This is based on the peer group consensus of all qualitative results returned from participants within the EQA challenge / distribution.

[3] **Sample Status:** EQA samples are defined as "CORE" or "EDUCATIONAL". Core proficiency samples are reviewed by the QCMD Scientific Expert(s). This is on the basis of scientific information, clinical relevance, current literature and, where appropriate, professional clinical guidelines. Participating laboratories are expected to report core proficiency samples correctly within the EQA challenge / distribution.

[4] **Consensus (IU/ml):** Mean consensus (Log<sub>10</sub>) calculated from data returned by participants with outliers removed and number of quantitative results (n) returned for each panel member.

[5] **Range:** Maximum and minimum quantitative value (IU/ml) reported by participants within this challenge without outliers removed.

*For further details please refer to the current participant manual.*

### Your Summary Results

<b>Units</b>	IU/ml
<b>EQA Assessment Group <sup>[1]</sup></b>	Commercial
<b>Core Panel Detection (Qualitative) Score <sup>[2]</sup></b>	<b>0</b>
<b>Core Panel Estimation (Quantitative) Score <sup>[3]</sup></b>	<b>1</b>

# Individual Report

# QCMD 2019 Human Herpes Virus 6 DNA EQA Programme



<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory:</b> CZ023
-------------------------------------	-------------------------------	-------------------------	---	---------------------------	--	-----------------------------

## Core Panel Members Results

Sample Code	Unitage	EQA Assessment Group Consensus [4]	SD [5]	Quantitative Result		Qualitative Result		
				Your Result [6]	Estimation Score [7]	Percentage Correct (All) [8]	Your Result [9]	Detection Score [10]
HHV6DNA19C1-01	IU/ml	4.446	0.335	4.813	1	100.0	Positive	0
HHV6DNA19C1-02	IU/ml	N/A	-	LOD/NR	N/A	100.0	Negative	0
HHV6DNA19C1-03	IU/ml	4.788	0.546	4.260	0	100.0	Positive	0
HHV6DNA19C1-05	IU/ml	4.163	0.403	4.050	0	100.0	Positive	0

All quantitative values above expressed in Log<sub>10</sub> IU/ml.

[1] **EQA Assessment Group:** To aid data analysis, participant results are grouped according to the molecular amplification/detection method specified within their molecular workflow for this challenge / distribution. For further details refer to the *Additional Information: Individual Panel Member Analysis* section of this report.

[2] **Core Panel Detection (Qualitative) Score:** An overall core panel detection score provided per challenge / distribution.

[3] **Core Panel Estimation (Quantitative) Score:** An overall core panel estimation score provided per challenge / distribution.

[4] **EQA Assessment Group Consensus:** The mean value for all results within your EQA assessment group.

[5] **SD:** The standard deviation for results from your EQA assessment group.

[6] **Your Quantitative Result:** The quantitative result you returned for each sample within this EQA challenge. LOD/NR (limit of detection or not reported).

[7] **Estimation Score:** Your estimation (quantitative) scores are calculated based on your variation from the consensus for your EQA assessment group. With 0 (zero) scored if the quantitative value you reported is within one standard deviation (SD) from your EQA assessment group consensus, 1 (one) if your quantitative value is between one and two SDs, 2 (two) if your quantitative value is within two and three SDs and 3 (three) if your quantitative value is more than three SDs from the mean of your EQA assessment group.

[8] **Percentage Correct (All):** Percentage of datasets (%) reporting the correct qualitative results for each panel member.



[9] **Your Qualitative Result:** The qualitative result you reported for each sample within this EQA challenge / distribution.

[10] **Detection Score:** Your detection (qualitative) scores are based on the assigned detection frequency of each panel members, where 0 (zero) is "highly satisfactory" and 3 (three) is "highly unsatisfactory". Scores are provided for individual panel members.

**For further details please refer to the current participant manual.**


<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 <small>Quality Control for Molecular Diagnostics</small>	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### Core Panel Member Score Breakdown




**Core Panel Member Score Breakdown - Detection:** This figure gives you a breakdown of the qualitative detection scores for all qualitative datasets returned within this EQA challenge / distribution independent of the EQA assessment group. Panel detection scores are generated from only those panel members that are defined as "CORE".

**Core Panel Member Score Breakdown - Estimation:** This figure gives you a breakdown of the quantitative estimation scores for all quantitative datasets returned within this EQA challenge / distribution independent of the EQA assessment group. Panel estimation scores are based on positive core panel members only. Those datasets that did not return quantitative values for all core samples are represented by '-'.  
*For further details please refer to the current participant manual.*

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 <small>Quality Control for Molecular Diagnostics</small>	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory:</b> CZ023

## Duplicate Sample Performance Over Time

### Series 4: HHV6 Type A



#### Observation Details

- 1: HHV6DNA18C2-05 - HHV6DNA18C2-01: Your Variation -1.062 (IU/ml): Overall Mean Variation -0.158: Overall SD 0.907.  
2: HHV6DNA19C1-01 - HHV6DNA18C2-05: Your Variation 0.758 (IU/ml): Overall Mean Variation 0.136: Overall SD 0.645.  
3: HHV6DNA19C1-05 - HHV6DNA19C1-01: Your Variation -0.763 (IU/ml): Overall Mean Variation -0.237: Overall SD 0.418.


QCMD monitors your laboratory's performance over time based on the reported quantitative variation between duplicate panel members within the EQA challenge and, where appropriate, across EQA challenges.

The mean variation and standard deviation are calculated from the quantitative variation reported by each participant between duplicate panel members in the same unit of measurement once outliers have been removed. (See 'Observation Details')

Previous and current observations are plotted on the chart as the number of standard deviations your variation was from the mean variation for all participants who submitted corresponding results in the same unitage.




Any reported variation greater than  $\pm 3$  SD will not be shown on the graph, but your variation value will be provided **in red** in the *Observation Details*.

When "N/A" is displayed for an observation, either no valid quantitative results were provided or there was a change in reported unitage.

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 <small>Quality Control for Molecular Diagnostics</small>	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory:</b> CZ023

## My Workflow Details

The details of the workflow(s) used to submit your results for this challenge.

<b>Name</b>	GeneProof Human Herpes Virus 6/7 (HHV-6/7) PCR kit (v2)
<b>Description</b>	
<b>Targets</b>	 human herpes virus
<b>Assays</b>	<ul style="list-style-type: none"> <li> <i>Extraction</i> - Manual Extraction Process <ul style="list-style-type: none"> <li>● Commercial <ul style="list-style-type: none"> <li>○ Kit Manufacturer: <i>GeneProof</i></li> <li>○ Kit Type: <i>PathogenFree DNA Isolation Kit</i></li> </ul> </li> </ul> </li> <li> <i>Amplification</i> - GeneProof - croBEE Real-Time PCR System <ul style="list-style-type: none"> <li>● Commercial <ul style="list-style-type: none"> <li>○ Kit Manufacturer: <i>GeneProof</i></li> <li>○ Kit Type: <i>Human Herpes Virus 6/7 (HHV-6/7) PCR Kit</i></li> <li>○ Kit Version: <i>ISEX</i></li> </ul> </li> </ul> </li> </ul>

## Educational Panel Members Results

Sample Code	Unitage	EQA Assessment Group Consensus <sup>[1]</sup>	SD <sup>[2]</sup>	Quantitative Result		Qualitative Result		
				Your Result <sup>[3]</sup>	Estimation Score <sup>[4]</sup>	Percentage Correct (All) <sup>[5]</sup>	Your Result <sup>[6]</sup>	Detection Score <sup>[7]</sup>
HHV6DNA19C1-04	IU/ml	2.930	0.392	2.320	1	96.8	Positive	0

All quantitative values above expressed in Log<sub>10</sub> IU/ml.

[1] **EQA Assessment Group Consensus:** The mean value for all results within your EQA assessment group.

[2] **SD:** The standard deviation for results from your EQA assessment group.

[3] **Your Quantitative Result:** The quantitative result you returned for each sample within this EQA challenge. LOD/NR (limit of detection or not reported).


[4] **Estimation Score:** Your estimation (quantitative) scores are calculated based on your variation from the consensus for your EQA assessment group. With 0 (zero) scored if the quantitative value you reported is within one standard deviation (SD) from your EQA assessment group consensus, 1 (one) if your quantitative value is between one and two SDs, 2 (two) if your quantitative value is within two and three SDs and 3 (three) if your quantitative value is more than three SDs from the mean of your EQA assessment group.

[5] **Percentage Correct (All):** Percentage of datasets (%) reporting the correct qualitative results for each panel member.

[6] **Your Qualitative Result:** The qualitative result you reported for each sample within this EQA challenge / distribution.

[7] **Detection Score:** Your detection (qualitative) scores are based on the assigned detection frequency of each panel members, where 0 (zero) is "highly satisfactory" and 3 (three) is "highly unsatisfactory". Scores are provided for individual panel members.

**For further details please refer to the current participant manual.**

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>				
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### Further Programme Details

Number of Participants	95
Number of Countries	23
Number of Respondents	90
Number of Datasets Submitted	105
Quantitative Results Returned (All)	81 (77.1%)
- Quantitative Results Returned (IU/ml)	18 (22.2%)
- Quantitative Results Returned (Copies/ml)	63 (77.8%)
Qualitative Results Returned	93 (88.6%)

### EQA Programme Aims


To assess the proficiency of laboratories' molecular assays in the detection of various types of human herpes virus 6 (HHV6).

To assess the proficiency of laboratories in the reliable quantitation of HHV6 viral load.

### Feedback and Enquiries

Participants are encouraged to read the QCMD Participants' Manual, which can be downloaded from the QCMD website.

Any queries about this report should be addressed to the QCMD Neutral Office ([neutraloffice@qcmd.org](mailto:neutraloffice@qcmd.org)).

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>				
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory:</b> CZ023

Panel member analysis is separated into CORE samples followed by EDUCATIONAL samples.

### Additional Core Samples Information

The following section has been categorised as shown below:

Core ► Quantitative ► IU/ml, Copies/ml ► Qualitative


### Individual Panel Member Analysis (Quantitative)

Quantitative analysis for each panel member is provided in relation to your EQA assessment group. EQA assessment groups are established using the molecular workflow information reported by all participants within this EQA challenge / distribution. The principal level of assessment is at the individual method level which is defined based on your reported “amplification/detection method” and other laboratories using the same or similar amplification/detection methods.

To allow meaningful assessment at the individual method level the EQA assessment group must consist of 5 or more datasets. If there are not sufficient datasets at the individual method level then your results will be included within a higher EQA assessment group based on whether it is a commercial or in house technology/method. The highest level assessment grouping is all reported results using the same unit of measurement (i.e. Copies/ml or IU/ml).

The results below provide a breakdown of participant reported values on each of the panel members within this EQA challenge / distribution. Your result for each panel member is indicated by "your value". You can compare your value to the “mean” within your EQA assessment group and the overall consensus for each sample within this EQA challenge / distribution.


#### Key



Catalogue Code: QAV084119	Ref Code: HHV6DNA19	Challenge: C1	Analysis Type: Qualitative and Quantitative	Dataset: 278248	Report UID: 2677/278248/2033	Laboratory CZ023
------------------------------	------------------------	------------------	--	--------------------	---------------------------------	---------------------

**HHV6DNA19C1-01 - Quantitative Results Breakdown (IU/ml)**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (IU/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-01	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	4.388	18	3.753 - 5.118




**Groups below n=5:** GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2), In-House (n=3), In-House - Real-time In-House PCR (n=3)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=11)

**HHV6DNA19C1-03 - Quantitative Results Breakdown (IU/ml)**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (IU/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-03	HHV6 Type A	Plasma		Frequently Detected	CORE	4.775	18	3.773 - 5.592



**Groups below n=5:** GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2), In-House (n=3), In-House - Real-time In-House PCR (n=3)


**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=11)



Catalogue Code: QAV084119	Ref Code: HHV6DNA19	Challenge: C1	Analysis Type: Qualitative and Quantitative	Dataset: 278248	Report UID: 2677/278248/2033	Laboratory CZ023
------------------------------	------------------------	------------------	--	--------------------	---------------------------------	---------------------

**HHV6DNA19C1-05 - Quantitative Results Breakdown (IU/ml)**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (IU/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-05	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	4.151	18	3.138 - 4.526




**Groups below n=5:** GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2), In-House (n=3), In-House - Real-time In-House PCR (n=3)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=11)

<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023
-------------------------------------	-------------------------------	-------------------------	---	---------------------------	--	----------------------------

**HHV6DNA19C1-01 - Quantitative Results Breakdown (Copies/ml)**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (Copies/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-01	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	4.054	63	3.264 - 5.106




**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), ELITech Group - Elitech Alert Real Time Q-PCR (n=1), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=8), bioMerieux - bioMerieux R-gene Quant Kit (n=12), In-House - Real-time In-House PCR (n=22)

<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023
-------------------------------------	-------------------------------	-------------------------	---	---------------------------	--	----------------------------

**HHV6DNA19C1-03 - Quantitative Results Breakdown (Copies/ml)**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (Copies/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-03	HHV6 Type A	Plasma		Frequently Detected	CORE	4.522	63	3.286 - 5.911




**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), ELITech Group - Elitech Alert Real Time Q-PCR (n=1), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=8), bioMerieux - bioMerieux R-gene Quant Kit (n=12), In-House - Real-time In-House PCR (n=22)

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 Quality Control for Molecular Diagnostics	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### HHV6DNA19C1-05 - Quantitative Results Breakdown (Copies/ml)

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (Copies/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-05	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	4.096	61	2.124 - 6.023



**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), ELITech Group - Elitech Alert Real Time Q-PCR (n=1), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=8), bioMerieux - bioMerieux R-gene Quant Kit (n=12), In-House - Real-time In-House PCR (n=20)

### Individual Panel Member Analysis (Qualitative)

Qualitative analysis for each panel member is provided in relation to your EQA assessment group. EQA assessment groups are established using the molecular workflow information reported by all participants within this EQA challenge / distribution. The principal level of assessment is at the individual method level which is defined based on your reported “amplification/detection method” and other laboratories using the same or similar amplification/detection methods.


To allow meaningful assessment at the individual method level the EQA assessment group must consist of 5 or more datasets. If there are not sufficient datasets at the individual method level then your results will be included within a higher EQA assessment group based on whether it is a commercial or in house technology/method. The highest level assessment grouping is “All” participant reported qualitative results.

A breakdown of qualitative results reported by participants on each of the panel members within this EQA challenge / distribution is provided below. You can compare your results to those within your EQA assessment group and those obtained within other EQA assessment groups or to the overall consensus for each sample within this EQA challenge / distribution.

Catalogue Code: QAV084119	Ref Code: HHV6DNA19	Challenge: C1	Analysis Type: Qualitative and Quantitative	Dataset: 278248	Report UID: 2677/278248/2033	Laboratory CZ023
------------------------------	------------------------	------------------	--	--------------------	---------------------------------	---------------------


**HHV6DNA19C1-01 - Qualitative Results Breakdown**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
HHV6DNA19C1-01	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	100.0	93




**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), Progenie Molecular (n=1), Progenie Molecular - Progenie Molecular RealCycler (n=1), QIAGEN (n=4), QIAGEN - QIAGEN Artus Real Time (n=4), Seegene (n=2), Seegene - Seegene Seplex (n=2), bioMerieux - BioFire FilmArray (n=1), fast-track DIAGNOSTICS (n=3), fast-track DIAGNOSTICS - FTD real time PCR (n=3), In-House - Conventional In-House PCR (n=1)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=15), ELITech Group - Elitech Elite Real Time kit (n=10)

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 <small>Quality Control for Molecular Diagnostics</small>	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### HHV6DNA19C1-02 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
HHV6DNA19C1-02	HHV6 Negative	Plasma		Negative	CORE	100.0	93




**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), Progenie Molecular (n=1), Progenie Molecular - Progenie Molecular RealCycler (n=1), QIAGEN (n=4), QIAGEN - QIAGEN Artus Real Time (n=4), Seegene (n=2), Seegene - Seegene Seplex (n=2), bioMerieux - BioFire FilmArray (n=1), fast-track DIAGNOSTICS (n=3), fast-track DIAGNOSTICS - FTD real time PCR (n=3), In-House - Conventional In-House PCR (n=1)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=15), ELITech Group - Elitech Elite Real Time kit (n=10)

Catalogue Code: QAV084119	Ref Code: HHV6DNA19	Challenge: C1	Analysis Type: Qualitative and Quantitative	Dataset: 278248	Report UID: 2677/278248/2033	Laboratory CZ023
------------------------------	------------------------	------------------	--	--------------------	---------------------------------	---------------------


**HHV6DNA19C1-03 - Qualitative Results Breakdown**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
HHV6DNA19C1-03	HHV6 Type A	Plasma		Frequently Detected	CORE	100.0	93




**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), Progenie Molecular (n=1), Progenie Molecular - Progenie Molecular RealCycler (n=1), QIAGEN (n=4), QIAGEN - QIAGEN Artus Real Time (n=4), Seegene (n=2), Seegene - Seegene Seplex (n=2), bioMerieux - BioFire FilmArray (n=1), fast-track DIAGNOSTICS (n=3), fast-track DIAGNOSTICS - FTD real time PCR (n=3), In-House - Conventional In-House PCR (n=1)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=15), ELITech Group - Elitech Elite Real Time kit (n=10)

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>				
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### HHV6DNA19C1-05 - Qualitative Results Breakdown


Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
HHV6DNA19C1-05	HHV6 Type A	Plasma	D1	Frequently Detected	CORE	100.0	93



**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), Progenie Molecular (n=1), Progenie Molecular - Progenie Molecular RealCycler (n=1), QIAGEN (n=4), QIAGEN - QIAGEN Artus Real Time (n=4), Seegene (n=2), Seegene - Seegene Seplex (n=2), bioMerieux - BioFire FilmArray (n=1), fast-track DIAGNOSTICS (n=3), fast-track DIAGNOSTICS - FTD real time PCR (n=3), In-House - Conventional In-House PCR (n=1)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=15), ELITech Group - Elitech Elite Real Time kit (n=10)



<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>				
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### Additional Educational Samples Information

The following section has been categorised as shown below:

Educational ► Quantitative ► IU/ml, Copies/ml ► Qualitative


### Individual Panel Member Analysis (Quantitative)

Quantitative analysis for each panel member is provided in relation to your EQA assessment group. EQA assessment groups are established using the molecular workflow information reported by all participants within this EQA challenge / distribution. The principal level of assessment is at the individual method level which is defined based on your reported “amplification/detection method” and other laboratories using the same or similar amplification/detection methods.

To allow meaningful assessment at the individual method level the EQA assessment group must consist of 5 or more datasets. If there are not sufficient datasets at the individual method level then your results will be included within a higher EQA assessment group based on whether it is a commercial or in house technology/method. The highest level assessment grouping is all reported results using the same unit of measurement (i.e. Copies/ml or IU/ml).

The results below provide a breakdown of participant reported values on each of the panel members within this EQA challenge / distribution. Your result for each panel member is indicated by "your value". You can compare your value to the “mean” within your EQA assessment group and the overall consensus for each sample within this EQA challenge / distribution.


#### Key



<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 Quality Control for Molecular Diagnostics	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

**HHV6DNA19C1-04 - Quantitative Results Breakdown (IU/ml)**

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (IU/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-04	HHV6 Type B	Plasma		Frequently Detected	EDUCATIONAL	2.901	17	2.107 - 3.518




**Groups below n=5:** GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2), In-House (n=3), In-House - Real-time In-House PCR (n=3)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=10)

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 Quality Control for Molecular Diagnostics	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### HHV6DNA19C1-04 - Quantitative Results Breakdown (Copies/ml)

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Consensus (Copies/ml)		Range
						(Log <sub>10</sub> )	(n)	
HHV6DNA19C1-04	HHV6 Type B	Plasma		Frequently Detected	EDUCATIONAL	2.900	56	1.982 - 4.228



**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), ELITech Group - Elitech Alert Real Time Q-PCR (n=1), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), QIAGEN (n=2), QIAGEN - QIAGEN Artus Real Time (n=2)


**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=8), bioMerieux - bioMerieux R-gene Quant Kit (n=10), In-House - Real-time In-House PCR (n=19)

### Individual Panel Member Analysis (Qualitative)

Qualitative analysis for each panel member is provided in relation to your EQA assessment group. EQA assessment groups are established using the molecular workflow information reported by all participants within this EQA challenge / distribution. The principal level of assessment is at the individual method level which is defined based on your reported “amplification/detection method” and other laboratories using the same or similar amplification/detection methods.


To allow meaningful assessment at the individual method level the EQA assessment group must consist of 5 or more datasets. If there are not sufficient datasets at the individual method level then your results will be included within a higher EQA assessment group based on whether it is a commercial or in house technology/method. The highest level assessment grouping is “All” participant reported qualitative results.

A breakdown of qualitative results reported by participants on each of the panel members within this EQA challenge / distribution is provided below. You can compare your results to those within your EQA assessment group and those obtained within other EQA assessment groups or to the overall consensus for each sample within this EQA challenge / distribution.

<b>Individual Report</b>		<b>QCMD 2019 Human Herpes Virus 6 DNA EQA Programme</b>			 Quality Control for Molecular Diagnostics	
<b>Catalogue Code:</b> QAV084119	<b>Ref Code:</b> HHV6DNA19	<b>Challenge:</b> C1	<b>Analysis Type:</b> Qualitative and Quantitative	<b>Dataset:</b> 278248	<b>Report UID:</b> 2677/278248/2033	<b>Laboratory</b> CZ023

### HHV6DNA19C1-04 - Qualitative Results Breakdown

Sample Code	Sample Content	Matrix	Sample Relationships	Detection Frequency	Sample Status	Percentage Correct (All)	
						(%)	(n)
HHV6DNA19C1-04	HHV6 Type B	Plasma		Frequently Detected	EDUCATIONAL	96.8	93



**Groups below n=5:** AB Analytica (n=4), AB Analytica - AB Analytica REALQUALITY RQ (n=4), Clonit (n=1), Clonit - Clonit PCR Reagents (n=1), GeneProof (n=2), GeneProof - GeneProof Real Time PCR kit (n=2), Mobius (n=1), Mobius - Mobius PCR Reagents (n=1), Progenie Molecular (n=1), Progenie Molecular - Progenie Molecular RealCycler (n=1), QIAGEN (n=4), QIAGEN - QIAGEN Artus Real Time (n=4), Seegene (n=2), Seegene - Seegene Seplex (n=2), bioMerieux - BioFire FilmArray (n=1), fast-track DIAGNOSTICS (n=3), fast-track DIAGNOSTICS - FTD real time PCR (n=3), In-House - Conventional In-House PCR (n=1)

**Groups Rolled Up:** Altona Diagnostics - Altona Diagnostics RealStar (n=15), ELITech Group - Elitech Elite Real Time kit (n=10)

**QCMD © 2019.** The QCMD EQA programme samples, associated reports and data generated during this programme are intended for External Quality Assessment (EQA) and Proficiency Testing (PT) purposes only. QCMD operates according to a strict Code of Practice which is in line with ISO/IEC 17043 and associated standards. Data reported in QCMD programmes is representative of a laboratory's standard diagnostic testing protocols irrespective of the technology they use. The data provided in the reports are based on technical information provided by the individual laboratories as part of the assessment process, as such it does not constitute a formal technology method comparison. All text and images produced by QCMD are the property of QCMD unless otherwise stated.

The reproduction and use of these materials is not permitted without the express written consent of QCMD. The use of the information provided in QCMD reports for commercial purposes is strictly prohibited.